



ANALIZA BIAŁEK



100007274

Laboratory of Analysis

Client: EV Pharm
Order number: 100007274
Name: EV PHARM DROSTANOLONE E-200
Number of samples: 1 vial

Batch number: M7119AE
Recive date: 2025-03-11
Start date: 2025-03-17
End date: 2025-03-18

Sample identification

Substance	Method of taking a sample	Form
Drostanolone Enanthate	Sample provided by customer	Transparent liquid in a glass vial


Results

Substance	Sample name	Content	Purity
Drostanolone Enanthate	EV PHARM DROSTANOLONE E-200 200MG/ML	195,48mg/ml	-----

Research comment

The test sample was applied in the amount of 20 μ l to the BioBasic 8 Thermo Scientific column. During the analysis with the method a linear gradient in the solutions of H₂O and Acetonitrile with the addition of 0.1% trifluoroacetic acid, a peak of the sought substance appeared. Its concentration, when compared to the reference chromatogram, is 195,48mg/ml. The measurement error is + - 1mg. The actual state of the sample sent before the test is presented on the photo below

Photo of the test sample



Password

R62MSK83


Analiza Białek sp. z o.o.
KRS: 0000966168
NIP: 8992922970
REGON: 521751875

Headquarters and Laboratory:
ul. Duńska 9, 54-427 Wrocław
Budynek DELTA

Lab technician:
Analysis and research specialist
mgr Arkadiusz Zajęc



Chromatogram



Method

Time	H2O + 0,1% TFA	ACN + 0,1% TFA
0min	60%	40%
2min	60%	40%
27min	5%	95%
30min	5%	95%

Password

R62MSK83